

LIST OF IDENTIFYING NUMBERS

Alphapharm Pty Ltd, following consultation with the Therapeutic Goods Administration, is recalling the following batches of EpiPen® 300 microgram (µg) Adrenaline Injection Syringe Auto-Injectors due to the potential that these devices may contain a defective part that may result in the device failing to activate or requiring increased force to activate.

Batch number	Expiry
5FA665	Apr 17
5FA6651	Apr 17
5FA6652	Apr 17
5FA6653	Apr 17