

# Product Safety Recall



## GasBadge® Plus Single Gas Monitor

### INSTRUMENTS

GasBadge® Plus – Carbon Monoxide  
GasBadge® Plus – Hydrogen Sulfide  
GasBadge® Plus – Oxygen  
GasBadge® Plus – Nitrogen Dioxide  
GasBadge® Plus – Sulfur Dioxide  
xx = optional alphanumeric designation

### PART NUMBERS

18100050-1, 18100050-1xx, IPLU501  
18100050-2, 18100050-2xx, IPLU502  
18100050-3, 18100050-3xx, IPLU503  
18100050-4, 18100050-4xx, IPLU504  
18100050-5, 18100050-5xx, IPLU505

All current instruments (v2.2 and earlier; manufactured prior to February 2012)

**Defect** May not properly indicate a low battery condition.

**Hazard** Failure to alarm or instrument shut down. This condition is exacerbated when the instrument is used in temperatures below room temperature (72°F / 22°C).

**What to do** All customers should contact Industrial Scientific to arrange for the exchange of the Recalled GasBadge Plus Instruments.

**Contact details** Customers should complete the recall form found at [www.indsci.com/gbplusrecall](http://www.indsci.com/gbplusrecall) or contact Industrial Scientific Pty. at +(03) 9644 7777.

See [www.recalls.gov.au](http://www.recalls.gov.au) for  
Australian Product Recall Information